
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004995
11:24 24/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:46
PRESSURIZING 1:51
STERILIZING 2:46
136.3 C 223kPa 2:46
136.3 C 224kPa 3:16
136.4 C 221kPa 3:46
136.3 C 224kPa 4:16
136.3 C 221kPa 4:46
136.4 C 222kPa 5:16
136.4 C 223kPa 5:46
136.4 C 225kPa 6:16
136.4 C 225kPa 6:16
VENTING 6:16
AIR DRYING 7:12
CYCLE COMPLETE 13:14

Digital Signature #
7B15327A144C0431
