
Midmark M11 - v1.0.0
Total Cycles: 3246

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2023 04:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 217.5 F 0.0
2:00 214.3 F 4.0
4:00 223.4 F 9.9
6:00 228.3 F 6.7
8:00 238.8 F 11.8
10:00 237.7 F 10.0
12:00 237.6 F 8.6
14:00 251.3 F 16.1
16:00 263.9 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.1
2:00 271.6 F 28.3
3:00 272.3 F 29.1
4:00 272.3 F 28.9

Min 271.2 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.1 F 0.0
5:00 224.8 F 0.0
10:00 238.7 F 0.0
15:00 235.1 F 0.0
20:00 226.4 F 0.0
25:00 224.5 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 2:47
HEATING: 17:59
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

05/16/2023 05:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____