
Midmark M11 - v1.0.0
Total Cycles: 5261

Name: RITTER 1
Sterilizer ID: V2178198
01/22/2025 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.8 F 0.0
2:00 132.1 F 1.9
4:00 152.4 F 5.0
6:00 175.8 F 8.7
8:00 202.7 F 12.4
10:00 213.1 F 4.8
12:00 225.7 F 8.8
14:00 236.4 F 13.0
16:00 234.8 F 9.2
18:00 241.7 F 10.7
20:00 245.0 F 12.7
22:00 256.3 F 18.8
24:00 266.8 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 29.2
2:00 272.4 F 28.8
3:00 271.5 F 28.3
4:00 271.7 F 28.9

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 206.0 F 0.0
10:00 210.4 F 0.0
15:00 197.2 F 0.0
20:00 199.2 F 0.0
25:00 208.7 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 25:39
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:05:03
CYCLE COMPLETE

01/22/2025 04:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____