
Midmark M11 - v1.0.5
Total Cycles: 2093

Name: RITTER2
Sterilizer ID: V2313228
02/27/2023 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.6 F 0.0
2:00 153.4 F 2.1
4:00 163.6 F 5.1
6:00 180.1 F 9.4
8:00 218.3 F 7.8
10:00 226.2 F 9.4
12:00 232.7 F 8.9
14:00 237.5 F 11.7
16:00 237.3 F 9.9
18:00 250.4 F 16.4
20:00 262.2 F 23.2
22:00 273.0 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.9
1:00 272.2 F 29.8
2:00 272.4 F 30.0
3:00 272.5 F 30.3
4:00 272.3 F 30.5

Min 271.8 F 29.8
Max 272.7 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 217.8 F 0.0
10:00 221.7 F 0.0
15:00 223.1 F 0.0
20:00 231.3 F 0.0
25:00 224.5 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:09
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

02/27/2023 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____