
Midmark M11 - v1.0.5
Total Cycles: 3513

Name: RITTER2
Sterilizer ID: V2313228
04/10/2024 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.3 F 0.0
2:00 98.3 F 0.1
4:00 114.1 F 0.5
6:00 148.9 F 3.5
8:00 185.9 F 7.4
10:00 202.5 F 12.3
12:00 215.0 F 5.6
14:00 227.2 F 10.1
16:00 232.7 F 8.7
18:00 241.0 F 12.2
20:00 239.8 F 10.8
22:00 252.9 F 17.3
24:00 264.5 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.3 F 29.4
2:00 272.1 F 29.7
3:00 271.9 F 29.1
4:00 272.5 F 29.9

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 206.4 F 0.0
10:00 197.9 F 0.0
15:00 198.3 F 0.0
20:00 196.4 F 0.0
25:00 194.3 F 0.0
30:00 192.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:55
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:04:08
CYCLE COMPLETE

04/10/2024 02:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____