
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003811
13:11 09/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:41
STERILIZING 3:34
136.0 C 220kPa 3:34
136.4 C 226kPa 4:05
136.4 C 225kPa 4:35
136.3 C 223kPa 5:05
136.4 C 220kPa 5:35
136.4 C 223kPa 6:05
136.4 C 225kPa 6:35
136.1 C 221kPa 7:05
136.1 C 221kPa 7:05
VENTING 7:05
AIR DRYING 7:51
CYCLE COMPLETE 7:58

Digital Signature #
8A613C0A809E820B
