
Midmark M11 - v1.0.0
Total Cycles: 2285

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2022 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.9 F 0.0
2:00 198.6 F 4.6
4:00 216.0 F 10.6
6:00 228.8 F 7.1
8:00 244.4 F 13.5
10:00 242.5 F 11.6
12:00 240.6 F 10.4
14:00 254.7 F 17.6
16:00 266.5 F 24.7
18:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 28.8
2:00 271.9 F 28.8
3:00 272.6 F 28.7
4:00 271.8 F 28.8

Min 271.5 F 28.3
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 235.8 F 0.0
10:00 221.0 F 0.0
15:00 237.7 F 0.0
20:00 223.4 F 0.0
25:00 229.2 F 0.0
30:00 232.4 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 18:16
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

09/02/2022 03:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____