
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000764
9:12 28/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:23
STERILIZING 4:29
136.0 C 228kPa 4:29
136.1 C 231kPa 4:59
136.1 C 230kPa 5:29
136.3 C 231kPa 5:59
136.4 C 232kPa 6:29
136.3 C 232kPa 6:59
136.3 C 230kPa 7:29
136.2 C 230kPa 7:59
136.2 C 230kPa 7:59
VENTING 7:59
AIR DRYING 8:57
CYCLE COMPLETE 9:02

Digital Signature #
394AA50E6AE3DD41
