
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003484
7:44 20/04/2022

WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 4:53
STERILIZING 6:00
136.4 C 220kPa 6:00
136.3 C 223kPa 6:30
136.4 C 220kPa 7:00
136.3 C 221kPa 7:30
136.3 C 220kPa 8:00
135.9 C 219kPa 8:30
136.4 C 219kPa 9:00
136.4 C 223kPa 9:30
136.2 C 218kPa 10:00
136.4 C 222kPa 10:30
136.1 C 221kPa 11:00
136.4 C 223kPa 11:30
136.2 C 222kPa 12:00
136.3 C 222kPa 12:30
CYCLE ABORTED 12:57
STOP BUTTON PRESSED
NOT STERILE
END TIME 12:57

Digital Signature #
0886395080902871
