

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1748

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2022 12:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
2:00 186.6 F 3.3  
4:00 197.6 F 8.3  
6:00 224.4 F 8.7  
8:00 232.4 F 9.0  
10:00 238.4 F 10.1  
12:00 243.4 F 12.6  
14:00 242.3 F 11.5  
16:00 255.6 F 18.4  
18:00 267.3 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 272.5 F 29.1  
2:00 271.2 F 28.4  
3:00 272.3 F 28.9  
4:00 271.3 F 28.6  
  
Min 271.2 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.2 F 0.0  
5:00 218.1 F 0.0  
10:00 201.4 F 0.0  
15:00 204.5 F 0.0  
20:00 203.7 F 0.0  
25:00 209.3 F 0.0  
30:00 216.0 F 0.0  
  
Drying Complete

FILLING: 1:59  
HEATING: 19:14  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:25  
CYCLE COMPLETE

03/25/2022 02:02 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02