
Midmark M11 - v1.0.0
Total Cycles: 3286

Name: RITTER 1
Sterilizer ID: V2178198
05/25/2023 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.8 F 0.0
2:00 201.4 F 3.4
4:00 219.5 F 9.3
6:00 229.0 F 7.1
8:00 239.2 F 11.6
10:00 238.2 F 9.8
12:00 238.1 F 8.5
14:00 252.0 F 16.1
16:00 264.7 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.3 F 28.8
2:00 272.5 F 28.6
3:00 271.5 F 28.0
4:00 272.5 F 28.9

Min 271.4 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.7 F 0.0
5:00 242.5 F 0.0
10:00 235.3 F 0.0
15:00 228.8 F 0.0
20:00 228.4 F 0.0
25:00 228.9 F 0.0
30:00 228.2 F 0.0

Drying Complete

FILLING: 2:26
HEATING: 17:41
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:20
CYCLE COMPLETE

05/25/2023 02:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____