

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4596

Name: RITTER 1  
Sterilizer ID: V2178198  
05/30/2024 09:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.9 F 0.0  
2:00 154.0 F 1.7  
4:00 154.0 F 4.6  
6:00 171.6 F 8.1  
8:00 191.7 F 12.0  
10:00 209.7 F 4.7  
12:00 220.6 F 9.0  
14:00 233.4 F 13.4  
16:00 234.7 F 9.6  
18:00 238.8 F 9.8  
20:00 244.9 F 12.9  
22:00 255.7 F 18.7  
24:00 265.9 F 24.8  
26:00 271.4 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 273.0 F 29.6  
2:00 272.0 F 29.0  
3:00 272.2 F 29.5  
4:00 272.7 F 29.4

Min 271.6 F 28.9  
Max 273.2 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.6 F 0.0  
5:00 210.9 F 0.0  
10:00 215.1 F 0.0  
15:00 231.6 F 0.0  
20:00 228.4 F 0.0  
25:00 227.9 F 0.0  
30:00 230.7 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 26:42  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:05:11  
CYCLE COMPLETE

05/30/2024 10:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_