
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004337
16:49 12/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:44
136.4 C 223kPa 3:44
136.3 C 224kPa 4:14
136.4 C 225kPa 4:44
136.4 C 224kPa 5:14
136.4 C 225kPa 5:44
136.4 C 225kPa 6:14
136.4 C 225kPa 6:44
136.4 C 225kPa 7:14
136.4 C 225kPa 7:14
VENTING 7:14
AIR DRYING 8:45
CYCLE COMPLETE 9:00

Digital Signature #
B275E0CD18B8E8FD
