
Midmark M11 - v1.0.0
Total Cycles: 1156

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2021 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 167.3 F 2.6
4:00 191.6 F 6.9
6:00 212.4 F 12.8
8:00 225.3 F 8.3
10:00 237.3 F 11.7
12:00 239.0 F 11.3
14:00 238.0 F 9.7
16:00 252.3 F 16.6
18:00 264.5 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.7 F 29.1
2:00 271.5 F 28.4
3:00 271.7 F 28.7
4:00 272.7 F 29.0

Min 271.0 F 28.1
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 232.5 F 0.0
10:00 224.1 F 0.0
15:00 222.0 F 0.0
20:00 225.3 F 0.0
25:00 229.1 F 0.0
30:00 229.5 F 0.0

Drying Complete

FILLING: 2:27
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

09/09/2021 12:44 PM
ACCEPTED: X
REJECTED:
Operator: U02