

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2689

Name: RITTER 1  
Sterilizer ID: V2178198  
12/19/2022 11:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 221.5 F 0.0  
2:00 208.9 F 2.6  
4:00 209.3 F 7.0  
6:00 223.5 F 12.6  
8:00 229.5 F 8.3  
10:00 240.2 F 11.2  
12:00 241.5 F 11.8  
14:00 240.4 F 10.5  
16:00 254.1 F 17.5  
18:00 266.3 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.2 F 28.9  
2:00 271.4 F 28.5  
3:00 272.2 F 28.9  
4:00 271.4 F 28.3  
  
Min 271.4 F 28.2  
Max 272.4 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 229.3 F 0.0  
10:00 228.5 F 0.0  
15:00 228.9 F 0.0  
20:00 233.2 F 0.0  
25:00 230.1 F 0.0  
30:00 227.4 F 0.0  
  
Drying Complete

FILLING: 1:56  
HEATING: 19:33  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

12/19/2022 11:58 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_