
Midmark M11 - v1.0.0
Total Cycles: 3155

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2023 03:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 179.1 F 0.0
4:00 204.4 F 1.1
6:00 219.6 F 6.2
8:00 233.1 F 11.3
10:00 233.5 F 8.1
12:00 242.1 F 11.3
14:00 245.9 F 13.1
16:00 245.3 F 12.6
18:00 259.1 F 20.0
20:00 271.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.5 F 28.4
2:00 271.8 F 28.3
3:00 272.5 F 28.4
4:00 271.5 F 28.1

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 221.3 F 0.0
10:00 242.4 F 0.0
15:00 237.1 F 0.0
20:00 230.3 F 0.0
25:00 222.3 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

04/24/2023 04:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____