

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002041  
8:39 16/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:30  
STERILIZING 3:31  
136.4 C 223kPa 3:31  
136.3 C 223kPa 4:01  
136.3 C 223kPa 4:31  
136.3 C 223kPa 5:01  
136.2 C 222kPa 5:31  
136.0 C 223kPa 6:01  
136.4 C 225kPa 6:31  
136.4 C 224kPa 7:01  
136.4 C 224kPa 7:01  
VENTING 7:02  
AIR DRYING 7:52  
CYCLE COMPLETE 7:57  
-----

Digital Signature #  
9FFC28ABF059A6D5  
-----