
Midmark M11 - v1.0.0
Total Cycles: 649

Name: MIDMARK
Sterilizer ID: V2178198
02/02/2021 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.6 F 0.0
2:00 134.8 F 0.6
4:00 161.2 F 3.0
6:00 181.2 F 7.1
8:00 205.5 F 12.1
10:00 216.8 F 5.3
12:00 232.4 F 10.4
14:00 231.1 F 7.5
16:00 243.5 F 13.2
18:00 242.4 F 11.4
20:00 255.8 F 18.2
22:00 267.2 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.7
1:00 272.0 F 28.6
2:00 271.6 F 28.7
3:00 271.6 F 28.8
4:00 271.6 F 28.9

Min 271.3 F 28.6
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.8 F 0.0
5:00 231.3 F 0.0
10:00 227.8 F 0.0
15:00 227.3 F 0.0
20:00 214.1 F 0.0
25:00 215.7 F 0.0
30:00 209.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:57
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:02:01
CYCLE COMPLETE

02/02/2021 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____