
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003937
10:17 05/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:43
STERILIZING 4:37
136.0 C 220kPa 4:37
136.1 C 222kPa 5:07
136.1 C 218kPa 5:37
136.4 C 225kPa 6:07
136.4 C 224kPa 6:37
136.4 C 224kPa 7:07
136.4 C 222kPa 7:37
136.4 C 222kPa 8:07
136.4 C 222kPa 8:07
VENTING 8:07
AIR DRYING 8:59
CYCLE COMPLETE 10:06

Digital Signature #
BC685E7B69979C6C
