
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000945
17:54 22/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:24
STERILIZING 4:54
136.4 C 232kPa 4:54
136.4 C 233kPa 5:24
136.4 C 231kPa 5:54
136.4 C 232kPa 6:24
136.3 C 232kPa 6:54
136.1 C 230kPa 7:24
136.4 C 230kPa 7:54
136.3 C 231kPa 8:24
136.3 C 231kPa 8:24
VENTING 8:24
AIR DRYING 9:19
CYCLE COMPLETE 24:19

Digital Signature #
4FA96B70CE380611
