
Midmark M11 - v1.0.5
Total Cycles: 2199

Name: RITTER2
Sterilizer ID: V2313228
03/30/2023 11:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.0 F 0.0
2:00 192.3 F 4.5
4:00 207.1 F 10.2
6:00 224.7 F 6.0
8:00 235.1 F 12.3
10:00 238.2 F 10.7
12:00 238.7 F 9.9
14:00 253.3 F 17.3
16:00 265.7 F 24.7
18:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.9 F 29.8
2:00 271.8 F 29.1
3:00 273.1 F 29.9
4:00 272.4 F 29.4

Min 271.8 F 29.1
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 225.1 F 0.0
10:00 239.0 F 0.0
15:00 229.0 F 0.0
20:00 230.1 F 0.0
25:00 227.1 F 0.0
30:00 228.2 F 0.0

Drying Complete

FILLING: 2:20
HEATING: 18:27
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

03/30/2023 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____