
Midmark M11 - v1.0.0
Total Cycles: 4155

Name: RITTER 1
Sterilizer ID: V2178198
12/22/2023 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 177.4 F 3.0
4:00 184.5 F 6.9
6:00 203.6 F 11.4
8:00 221.3 F 4.7
10:00 228.1 F 9.3
12:00 238.4 F 10.3
14:00 241.2 F 11.2
16:00 238.4 F 9.4
18:00 251.4 F 15.6
20:00 262.0 F 21.7
22:00 271.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.6 F 28.6
2:00 272.5 F 28.8
3:00 271.5 F 28.5
4:00 271.8 F 28.4

Min 271.2 F 28.2
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 188.0 F 0.0
10:00 195.4 F 0.0
15:00 210.4 F 0.0
20:00 225.0 F 0.0
25:00 231.0 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:20
CYCLE COMPLETE

12/22/2023 03:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____