

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4388

Name: RITTER 1  
Sterilizer ID: V2178198  
03/20/2024 04:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.2 F 0.0  
2:00 105.9 F 1.0  
4:00 127.0 F 3.4  
6:00 158.0 F 6.7  
8:00 182.8 F 10.8  
10:00 205.0 F 3.9  
12:00 216.3 F 7.5  
14:00 228.3 F 11.8  
16:00 230.5 F 7.7  
18:00 242.9 F 13.1  
20:00 241.1 F 11.2  
22:00 253.6 F 17.5  
24:00 264.8 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.4 F 29.1  
2:00 272.0 F 29.1  
3:00 272.0 F 28.8  
4:00 272.3 F 29.4

Min 271.4 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 221.9 F 0.0  
10:00 218.9 F 0.0  
15:00 216.6 F 0.0  
20:00 226.1 F 0.0  
25:00 224.4 F 0.0  
30:00 219.0 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 25:52  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:04:22  
CYCLE COMPLETE

03/20/2024 05:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_