
Midmark M11 - v1.0.5
Total Cycles: 4400

Name: RITTER2
Sterilizer ID: V2313228
03/05/2025 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.5 F 0.0
2:00 177.7 F 1.4
4:00 176.7 F 4.6
6:00 189.6 F 8.8
8:00 210.2 F 11.1
10:00 220.9 F 7.1
12:00 231.5 F 11.5
14:00 232.0 F 8.3
16:00 242.6 F 12.9
18:00 242.3 F 12.3
20:00 253.4 F 18.1
22:00 263.6 F 24.3
24:00 272.8 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 272.4 F 30.3
2:00 271.8 F 29.7
3:00 271.2 F 30.0
4:00 272.1 F 30.2

Min 271.0 F 29.5
Max 272.6 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 219.7 F 0.0
10:00 220.1 F 0.0
15:00 212.0 F 0.0
20:00 211.8 F 0.0
25:00 211.2 F 0.0
30:00 206.5 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 24:17
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

03/05/2025 12:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____