
Midmark M11 - v1.0.5
Total Cycles: 1083

Name: RITTER2
Sterilizer ID: V2313228
05/09/2022 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.4 F 0.0
2:00 163.8 F 3.1
4:00 179.5 F 7.1
6:00 201.6 F 12.0
8:00 219.0 F 6.3
10:00 232.9 F 11.4
12:00 232.9 F 8.7
14:00 239.6 F 10.4
16:00 247.0 F 13.9
18:00 258.6 F 20.5
20:00 269.6 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.3
2:00 271.9 F 29.5
3:00 271.9 F 29.0
4:00 272.6 F 29.7

Min 271.3 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 209.3 F 0.0
10:00 185.1 F 0.0
15:00 193.8 F 0.0
20:00 195.4 F 0.0
25:00 196.2 F 0.0
30:00 182.7 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

05/09/2022 11:59 AM
ACCEPTED: X
REJECTED:
Operator: U02