

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1072

Name: RITTER2  
Sterilizer ID: V2313228  
05/04/2022 01:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.2 F 0.0  
2:00 132.1 F 1.7  
4:00 148.9 F 4.7  
6:00 176.8 F 8.8  
8:00 195.9 F 10.4  
10:00 215.3 F 7.3  
12:00 228.9 F 12.0  
14:00 231.7 F 8.8  
16:00 239.9 F 10.8  
18:00 244.5 F 13.3  
20:00 256.5 F 19.7  
22:00 267.6 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.7 F 30.1  
2:00 271.7 F 29.3  
3:00 272.2 F 30.2  
4:00 272.4 F 29.8

Min 271.4 F 29.2  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.6 F 0.0  
5:00 200.1 F 0.0  
10:00 210.1 F 0.0  
15:00 213.2 F 0.0  
20:00 214.9 F 0.0  
25:00 215.6 F 0.0  
30:00 215.2 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 23:25  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:43  
CYCLE COMPLETE

05/04/2022 02:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_