
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004506
10:29 06/10/2022
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 4:52
STERILIZING 5:38
136.0 C 220kPa 5:38
136.4 C 225kPa 6:08
136.4 C 223kPa 6:38
135.9 C 220kPa 7:08
136.4 C 224kPa 7:38
136.4 C 222kPa 8:08
136.4 C 222kPa 8:38
136.4 C 224kPa 9:08
136.4 C 223kPa 9:38
136.3 C 224kPa 10:08
136.4 C 223kPa 10:38
136.2 C 223kPa 11:08
CYCLE ABORTED 11:36
STOP BUTTON PRESSED
NOT STERILE
END TIME 11:36

Digital Signature #
275EB0F22DBAD818
