
Midmark M11 - v1.0.5
Total Cycles: 4005

Name: RITTER2
Sterilizer ID: V2313228
10/07/2024 10:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.4 F 0.0
2:00 151.9 F 1.3
4:00 153.1 F 4.2
6:00 171.5 F 7.7
8:00 192.1 F 11.5
10:00 211.3 F 4.4
12:00 217.6 F 8.6
14:00 229.0 F 12.9
16:00 233.8 F 9.6
18:00 237.9 F 10.0
20:00 244.8 F 13.5
22:00 255.9 F 19.4
24:00 266.0 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 272.6 F 30.2
2:00 271.8 F 29.5
3:00 271.7 F 30.0
4:00 272.6 F 30.1

Min 271.1 F 29.3
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 196.6 F 0.0
10:00 191.1 F 0.0
15:00 187.6 F 0.0
20:00 180.3 F 0.0
25:00 172.1 F 0.0
30:00 170.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:49
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

10/07/2024 11:35 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____