
Midmark M11 - v1.0.5
Total Cycles: 146

Name: RITTER2
Sterilizer ID: V2313228
06/28/2021 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.2 F 0.0
2:00 181.6 F 1.8
4:00 180.1 F 5.4
6:00 195.0 F 9.9
8:00 214.7 F 7.5
10:00 226.5 F 9.9
12:00 233.3 F 9.1
14:00 241.2 F 12.6
16:00 241.2 F 11.5
18:00 252.9 F 17.5
20:00 263.4 F 23.7
22:00 272.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 271.9 F 29.8
2:00 271.7 F 29.2
3:00 271.4 F 29.5
4:00 272.3 F 29.6

Min 271.0 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 223.8 F 0.0
10:00 231.8 F 0.0
15:00 229.2 F 0.0
20:00 211.3 F 0.0
25:00 188.7 F 0.0
30:00 177.9 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

06/28/2021 09:46 AM
ACCEPTED: X
REJECTED:
Operator: U01