
Midmark M11 - v1.0.0
Total Cycles: 3641

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2023 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.5 F 0.0
2:00 165.5 F 2.1
4:00 164.6 F 5.4
6:00 185.6 F 9.4
8:00 211.7 F 6.6
10:00 222.0 F 7.7
12:00 231.9 F 12.8
14:00 235.3 F 9.4
16:00 238.8 F 9.6
18:00 246.1 F 13.5
20:00 257.9 F 19.9
22:00 268.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.2 F 28.9
2:00 272.6 F 29.4
3:00 271.5 F 28.6
4:00 272.3 F 29.0

Min 271.5 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.4 F 0.0
5:00 220.9 F 0.0
10:00 230.2 F 0.0
15:00 228.4 F 0.0
20:00 229.2 F 0.0
25:00 230.1 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:07
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

08/14/2023 10:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____