

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000150  
13:21 14/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:33  
PRESSURIZING 4:34  
STERILIZING 5:41  
136.0 C 222kPa 5:41  
136.2 C 224kPa 6:11  
136.3 C 223kPa 6:41  
136.1 C 223kPa 7:11  
136.1 C 225kPa 7:41  
136.2 C 221kPa 8:11  
136.3 C 225kPa 8:41  
136.3 C 225kPa 9:11  
136.3 C 225kPa 9:11  
VENTING 9:11  
AIR DRYING 10:13  
CYCLE COMPLETE 11:33  
-----

Digital Signature #  
6FF6393341FD2F92  
-----