
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000526
7:56 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:35
PRESSURIZING 4:35
STERILIZING 6:01
136.2 C 220kPa 6:01
136.2 C 219kPa 6:31
136.2 C 222kPa 7:01
136.2 C 221kPa 7:31
136.4 C 223kPa 8:01
136.4 C 222kPa 8:31
136.3 C 223kPa 9:01
136.2 C 223kPa 9:31
136.2 C 223kPa 9:31
VENTING 9:31
AIR DRYING 10:30
CYCLE COMPLETE 11:17

Digital Signature #
AA46FD7C4668B187
