

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2207

Name: RITTER 1  
Sterilizer ID: V2178198  
08/08/2022 12:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.4 F 0.0  
2:00 122.6 F 0.2  
4:00 148.5 F 2.8  
6:00 174.9 F 6.5  
8:00 199.5 F 10.4  
10:00 215.0 F 5.5  
12:00 226.4 F 8.1  
14:00 236.0 F 12.5  
16:00 233.7 F 8.7  
18:00 243.6 F 11.9  
20:00 242.9 F 11.9  
22:00 254.6 F 17.9  
24:00 265.2 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.6 F 29.1  
2:00 271.7 F 28.4  
3:00 272.4 F 29.2  
4:00 272.3 F 28.8

Min 271.5 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 217.5 F 0.0  
10:00 229.9 F 0.0  
15:00 218.5 F 0.0  
20:00 224.3 F 0.0  
25:00 220.2 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 25:55  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:03:57  
CYCLE COMPLETE

08/08/2022 01:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_