
Midmark M11 - v1.0.0
Total Cycles: 4476

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2024 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.7 F 0.0
2:00 129.0 F 1.3
4:00 133.0 F 3.8
6:00 159.6 F 7.1
8:00 191.0 F 11.3
10:00 210.1 F 4.1
12:00 223.3 F 8.5
14:00 234.7 F 13.1
16:00 235.0 F 9.4
18:00 237.9 F 9.5
20:00 245.9 F 13.2
22:00 257.6 F 19.4
24:00 268.2 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.4 F 28.7
2:00 271.6 F 28.8
3:00 272.2 F 29.0
4:00 272.3 F 28.9

Min 270.9 F 28.1
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 215.4 F 0.0
10:00 225.7 F 0.0
15:00 233.5 F 0.0
20:00 225.8 F 0.0
25:00 223.0 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:15
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:03:32
CYCLE COMPLETE

04/22/2024 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____