
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001340
14:36 10/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:56
STERILIZING 5:06
136.0 C 221kPa 5:06
136.1 C 221kPa 5:36
136.0 C 221kPa 6:06
136.3 C 223kPa 6:36
136.4 C 225kPa 7:06
136.3 C 224kPa 7:36
136.4 C 223kPa 8:06
136.3 C 224kPa 8:36
136.3 C 224kPa 8:36
VENTING 8:36
AIR DRYING 9:32
CYCLE COMPLETE 11:46

Digital Signature #
033E11B9093BA372
