
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001195
10:38 28/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:47
STERILIZING 4:00
136.4 C 223kPa 4:00
136.2 C 223kPa 4:30
136.3 C 222kPa 5:00
135.9 C 221kPa 5:30
136.4 C 224kPa 6:00
136.2 C 222kPa 6:30
136.4 C 225kPa 7:00
136.4 C 225kPa 7:30
136.4 C 225kPa 7:30
VENTING 7:30
AIR DRYING 8:22
CYCLE COMPLETE 10:43

Digital Signature #
347136D260E1199F
