
Midmark M11 - v1.0.5
Total Cycles: 4258

Name: RITTER2
Sterilizer ID: V2313228
01/10/2025 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.0 F 0.0
2:00 139.3 F 0.1
4:00 162.5 F 0.4
6:00 189.2 F 2.3
8:00 210.0 F 6.7
10:00 225.0 F 11.8
12:00 232.2 F 8.1
14:00 242.4 F 12.4
16:00 243.4 F 12.4
18:00 242.5 F 11.8
20:00 255.9 F 18.7
22:00 267.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 271.8 F 28.8
2:00 272.2 F 29.2
3:00 272.3 F 29.5
4:00 271.5 F 28.6

Min 271.4 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 250.4 F 0.0
10:00 226.1 F 0.0
15:00 189.5 F 0.0
20:00 180.2 F 0.0
25:00 178.9 F 0.0
30:00 178.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:18
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:01:54
CYCLE COMPLETE

01/10/2025 05:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____