
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003558
11:53 28/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:22
STERILIZING 3:16
136.0 C 220kPa 3:16
136.4 C 223kPa 3:46
136.2 C 223kPa 4:16
135.8 C 217kPa 4:46
136.3 C 225kPa 5:16
136.2 C 224kPa 5:46
136.2 C 219kPa 6:16
136.4 C 222kPa 6:46
136.4 C 222kPa 6:46
VENTING 6:46
AIR DRYING 7:42
CYCLE COMPLETE 7:51

Digital Signature #
DB2028B91FBBB55A
