
Midmark M11 - v1.0.0
Total Cycles: 3752

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2023 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.4 F 0.0
2:00 172.5 F 2.3
4:00 192.6 F 6.0
6:00 209.9 F 10.3
8:00 220.5 F 6.7
10:00 227.8 F 9.5
12:00 234.3 F 9.5
14:00 241.3 F 12.0
16:00 240.1 F 10.4
18:00 252.3 F 16.6
20:00 263.0 F 22.9
22:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.2 F 29.0
2:00 271.8 F 28.9
3:00 272.1 F 29.1
4:00 272.5 F 29.4

Min 271.0 F 28.3
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 217.1 F 0.0
10:00 226.3 F 0.0
15:00 228.3 F 0.0
20:00 231.5 F 0.0
25:00 228.9 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:01:41
CYCLE COMPLETE

09/11/2023 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____