
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002515
8:58 19/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:31
STERILIZING 3:44
136.4 C 223kPa 3:44
136.4 C 226kPa 4:14
135.9 C 221kPa 4:44
136.4 C 224kPa 5:14
136.1 C 224kPa 5:44
136.3 C 224kPa 6:14
136.4 C 225kPa 6:44
136.1 C 221kPa 7:14
136.1 C 221kPa 7:14
VENTING 7:14
AIR DRYING 8:11
CYCLE COMPLETE 38:11

Digital Signature #
AC7B5041DB9CFF45
