
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002214
8:50 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:29
STERILIZING 4:44
136.0 C 220kPa 4:44
136.4 C 222kPa 5:14
136.2 C 223kPa 5:44
136.4 C 224kPa 6:14
136.0 C 220kPa 6:44
136.4 C 224kPa 7:14
136.4 C 224kPa 7:44
136.1 C 223kPa 8:14
136.1 C 223kPa 8:14
VENTING 8:14
AIR DRYING 9:06
CYCLE COMPLETE 39:06

Digital Signature #
CB237ABDEFCE625D
