

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 003426  
8:11 12/04/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 2:48  
STERILIZING 3:54  
136.0 C 220kPa 3:54  
135.8 C 220kPa 4:25  
136.2 C 219kPa 4:55  
136.5 C 221kPa 5:25  
136.2 C 223kPa 5:55  
136.3 C 223kPa 6:25  
136.2 C 220kPa 6:55  
136.1 C 219kPa 7:25  
136.1 C 219kPa 7:25  
VENTING 7:25  
AIR DRYING 8:12  
CYCLE COMPLETE 8:17  
-----

Digital Signature #  
BC8B7BA9136F7163  
-----