

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2864

Name: RITTER 1  
Sterilizer ID: V2178198  
02/06/2023 05:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.6 F 0.0  
2:00 174.4 F 3.7  
4:00 188.9 F 8.0  
6:00 210.4 F 13.3  
8:00 224.5 F 8.1  
10:00 237.0 F 13.4  
12:00 237.9 F 10.5  
14:00 235.8 F 8.6  
16:00 250.5 F 15.6  
18:00 262.7 F 22.6  
20:00 272.0 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.2 F 28.8  
2:00 271.6 F 28.8  
3:00 271.9 F 28.7  
4:00 271.6 F 28.9  
  
Min 271.2 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 220.9 F 0.0  
10:00 247.1 F 0.0  
15:00 232.4 F 0.0  
20:00 224.0 F 0.0  
25:00 223.3 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 20:10  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:03  
CYCLE COMPLETE

02/06/2023 06:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_