
Midmark M11 - v1.0.5
Total Cycles: 2993

Name: RITTER2
Sterilizer ID: V2313228
10/11/2023 07:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 73.7 F 0.0
4:00 127.5 F 0.3
6:00 166.2 F 2.0
8:00 159.1 F 4.8
10:00 191.5 F 8.3
12:00 191.3 F 12.5
14:00 210.6 F 4.8
16:00 222.6 F 8.5
18:00 235.7 F 12.8
20:00 233.9 F 8.9
22:00 240.2 F 11.1
24:00 244.1 F 12.7
26:00 256.3 F 19.0
28:00 266.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.1 F 29.2
2:00 272.1 F 29.6
3:00 271.3 F 28.8
4:00 272.0 F 29.5

Min 271.1 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 224.3 F 0.0
10:00 214.6 F 0.0
15:00 219.3 F 0.0
20:00 223.8 F 0.0
25:00 220.3 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:07:31
CYCLE COMPLETE

10/11/2023 09:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____