
Midmark M11 - v1.0.0
Total Cycles: 3934

Name: RITTER 1
Sterilizer ID: V2178198
10/23/2023 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 166.5 F 1.5
4:00 163.5 F 4.4
6:00 179.7 F 7.6
8:00 200.5 F 12.2
10:00 216.6 F 5.1
12:00 222.3 F 9.4
14:00 239.0 F 10.8
16:00 237.5 F 10.5
18:00 237.9 F 8.4
20:00 247.6 F 14.1
22:00 258.4 F 19.9
24:00 268.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.7 F 29.2
2:00 272.6 F 28.9
3:00 271.8 F 28.5
4:00 271.1 F 28.4

Min 271.1 F 28.3
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 228.8 F 0.0
10:00 229.3 F 0.0
15:00 228.1 F 0.0
20:00 229.0 F 0.0
25:00 225.0 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 25:22
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:22
CYCLE COMPLETE

10/23/2023 10:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____