
Midmark M11 - v1.0.0
Total Cycles: 2122

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2022 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 183.1 F 2.5
4:00 196.2 F 7.0
6:00 216.5 F 12.1
8:00 225.8 F 6.7
10:00 238.4 F 11.9
12:00 235.4 F 8.9
14:00 239.2 F 9.6
16:00 247.3 F 13.8
18:00 259.7 F 20.6
20:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.2 F 28.7
2:00 271.7 F 28.7
3:00 272.1 F 28.5
4:00 271.9 F 28.9

Min 271.1 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 239.4 F 0.0
10:00 216.3 F 0.0
15:00 215.1 F 0.0
20:00 228.8 F 0.0
25:00 232.8 F 0.0
30:00 233.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

07/13/2022 02:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____