
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000898
15:35 02/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:56
CYCLE ABORTED 3:59
STOP BUTTON PRESSED
NOT STERILE
END TIME 3:59

Digital Signature #
1E661107007A9D77
