

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001201  
11:05 29/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:46  
PRESSURIZING 2:56  
STERILIZING 3:58  
136.0 C 220kPa 3:58  
136.4 C 226kPa 4:28  
136.4 C 225kPa 4:58  
136.4 C 224kPa 5:28  
136.3 C 225kPa 5:58  
136.2 C 222kPa 6:28  
136.1 C 222kPa 6:58  
136.4 C 225kPa 7:28  
136.4 C 225kPa 7:28  
VENTING 7:28  
AIR DRYING 8:16  
CYCLE COMPLETE 9:50  
-----

Digital Signature #  
01DC534B0563BFE2  
-----