
Midmark M11 - v1.0.5
Total Cycles: 750

Name: RITTER2
Sterilizer ID: V2313228
02/02/2022 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.4 F 0.0
2:00 148.2 F 1.1
4:00 162.4 F 4.4
6:00 184.1 F 8.8
8:00 211.2 F 8.4
10:00 224.7 F 8.4
12:00 234.8 F 11.8
14:00 238.9 F 11.3
16:00 238.1 F 9.9
18:00 251.9 F 16.6
20:00 263.4 F 23.5
22:00 271.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.8
2:00 272.3 F 29.3
3:00 271.2 F 28.8
4:00 271.7 F 29.5

Min 271.1 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 234.0 F 0.0
10:00 216.3 F 0.0
15:00 218.6 F 0.0
20:00 211.1 F 0.0
25:00 220.3 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

02/02/2022 10:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____