
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000699
11:23 16/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:39
STERILIZING 3:55
136.2 C 231kPa 3:55
136.4 C 233kPa 4:25
136.2 C 232kPa 4:55
136.1 C 232kPa 5:25
136.4 C 233kPa 5:55
136.4 C 233kPa 6:25
136.4 C 229kPa 6:55
136.4 C 233kPa 7:25
136.4 C 233kPa 7:25
VENTING 7:25
AIR DRYING 8:18
CYCLE COMPLETE 8:29

Digital Signature #
7433898DAA3A1EBB
